



Quality control laboratory

Product name	Buffer pH10	Batch no.	BPH100924
MFG .Date	09/2024	EXP.Date	09/2027
Sensitive	Air	Storage	RT

CAS no.	M.W	g/mol	M.F
---------	-----	-------	-----

S.N	Characteristics	Unite	Specification	Result
1	Appearance	---	Light violet	pass
2	Density	g/ml	1.01	pass
3	pH	At (20) C	10(±0.15)	pass

Above batch material complies with prescribed quality of the standard as stated above.

A.SATD

QC Manger

